

## **Description :**

LP6 Ventilator LP6 Ventilator are Mode: Standby, Assist/Control, SIMV or Pressure Cycle LP6 Ventilator are Low Pressure Alarm: Set limit for Low Pressure alarm 2 to 32 LP6 Ventilator are cmH2O in increments of 2 cmH2O LP6 Ventilator are High Pressure Alarm Limit: (Assist/Control, SIMV Modes) or limit LP6 Ventilator are for Pressure Cycle mode: 15 to 90 cmH2O in increments of 5 cmH2O LP6 Ventilator are Volume: Push-to-turn knob: 100 to 2200 mL LP6 Ventilator are Breath Rate: 1 to 20 BPM in increments of 1 BPM and 22 to 38 BPM LP6 Ventilator are in increments of 2 BPM LP6 Ventilator are Inspiratory Time: 0.5 to1.0 sec in increments of 0.1 sec; LP6 Ventilator are 1.2 sec; and 1.5 to 5.5 sec in increments of 0.5 sec LP6 Ventilator are Breathing Effort: Push-to-turn knob sets pressure level to detect LP6 Ventilator are breathing effort (-10 cmH2O/hPa to +10 cmH2O)

